

Doc: XF212CoA C732  
 Date: 14-April-2022  
 Rev: 1



## XF212 XerumFree™ Medium Supplement

# Certificate of Analysis

Description: XerumFree™, medium supplement  
 Catalog number: XF212  
 Lot number: **C732**  
 Expiration date: 4-April-2023  
 Storage temperature: 2°C to 8°C  
 Storage instructions: Store in the dark

TEST PARAMETERS	SPECIFICATION	RESULT	UNITS
<b>Appearance</b>	Light opalescent liquid	Passes	-
<b>Content</b>	100 ml/bottle	Passes	ml
	500 ml/bottle	Passes	ml
<b>Endotoxins</b>	< 7,5	0,06	IU/ml
<b>pH</b>	7 – 8	7,3	-
<b>Sterility testing</b>	no growth	Passes	-
<b>Osmolality</b>	285 - 325	306	mOsm/kg
<b>Mycoplasma</b>	<0,9	0,4	

Approved by:

M. van Tiel

Date: 14-April-2022

XerumFree is for in vitro laboratory use or further manufacturing only..